

AI-508 ARTIFICIAL INTELLIGENCE TEMPERATURE CONTROLLER

Operation Instruction Version 7.0

GENERAL

AI-508, an economical and easy operated model of Yudian artificial intelligent industry controller, is specially designed for use in the industries of plastic machinery, food machinery, packaging machinery, baking machinery and environmental testing equipment. It features:

- Application of artificial intelligence control algorithm with auto tuning function, avoiding the overshoot problem of standard PID algorithm and achieving precise control.
- Multiple thermocouples and RTDs are selectable. Integrated non-linear graduation tables, digital calibration and auto zero technology, and achieving accurate and stable measurement.
- Advanced modular structure, conveniently providing various outputs options, and making quick delivery and easy maintenance.
- universal 100~240VAC power supply.
- ISO9001, ISO14001 and CE certified, achieving world class level of quality, anti-interference ability and safety.

MODEL CODE SYMBOL

The type of AI-508T is made up of 5 parts, for example:

AI-508 A G L2 L2
① ② ③ ④ ⑤

① **shows the basal function of instrument**

AI-508 economical artificial intelligence temperature controller

② **shows the front panel dimension :**

A Front panel 96×96mm(width×height), cut out 92×92mm, depth behind mounting surface 100mm

D Front panel 72×72mm(width×height), cut out 68×68mm, depth behind mounting surface 95mm

D2 Front panel 48×48mm(width×height), cut out 45×45mm, depth behind mounting surface 95mm

E Front panel 48×96mm(width×height), cut out 45×92mm, depth behind mounting surface 100mm

F Front panel 96×48mm(width×height), cut out 92×45mm, depth behind mounting surface 100mm

③ **shows the module types of main output. Selectable modules are as follows:**

L1 Relay output module (Capacity: 2A/250VAC, normally open terminal can absorb electrical spark)

L4 Relay output module (Capacity: 2A/250VAC, small volume, normally open terminal can absorb electrical spark)

G SSR voltage output module (30mA/12VDC)

W1 TRIAC no contact normally open discrete output module, suitable for AC contactors up to 80A, and has low interference and long life.

K1 Single phase thyristor zero crossing trigger output module. (can trigger three-phase of TRIAC, a pair of inversely parallel connected SCRs or

TRIAC power module of current 5-500A)

K3 Three phases Thyristor zero crossing trigger output module. (can trigger three-phase of TRIAC, a pair of inversely parallel connected SCRs

or TRIAC power module of current 5-500A)

④ **shows the module type of alarm output. Selectable modules are as follows:**

N (or none) no module installed

L1 Relay contact output module (Capacity: 2A/250VAC, normal open terminal can absorb spark, support AL1 alarm output)

L2 Relay contact output module (Capacity: 1A/250VAC, small volume, support AL1 alarm output)

L5 Output module of dual normal open relay output (Capacity: 2A/250VAC, support AL1 and AL2 alarm output)

⑤ **shows the module type of Auxiliary output (AUX). Selectable modules are as follows:**

N (or none) no module installed

L1 Relay contact output module (Capacity: 2A/250VAC, normal open terminal can absorb spark, support AU1 alarm output)

L2 Relay contact output module (Capacity: 1A/250VAC, small volume, support AU1 alarm output)

L5 Output module of dual normal open relay outputs(Capacity: 2A/250VAC, support AU1 and AU2 alarm output)

Note1: For instrument of dimension D2, because of its limited volume, when L1 or L5 module is installed in the ALM socket, L1 can't be installed in the OUP socket, but L4, which is smaller, can be installed instead.

Note2: K3 can't be installed in instruments with dimension D or D2. There is no ALM socket in D2 instruments. L5 module can't be installed in the ALM socket of instrument with dimension D.


TECHNICAL SPECIFICATION

- Input type : K, S, R, E, J, N, Pt100
- Measurement range : K(0 ~1300℃), S(0 ~1700℃), R(0 ~1650℃), E(0 ~800℃), J(0 ~1000℃), N(0 ~1300℃), Pt100(-200 ~600℃)
- Measurement accuracy : 0.3%FS ±1℃
- Temperature display resolution : 1℃
- Control mode : On-off control mode or Artificial Intelligence control with auto tuning.
- Output mode (modularized) : L1, G, W1, K1, K3 modules are selectable
- Alarm function : High limit and deviation high alarm.
- Power supply voltage rating : 100-240VAC, -15%, +10%; 50-60Hz.
- Power consumption : ≤3W.
- Ambient : Temperature of -10 ~60℃, humidity of 0 ~90RH%

FRONT PANEL AND OPERATION

- ① Upper display window, displays PV, or code of a parameter
- ② Lower display window, displays SV, alarming code, or value of a parameter
- ③ Setup key, for accessing parameter table, and confirming change.
- ④ Data shift key, also for activating auto turning
- ⑤ Data decrease key
- ⑥ Data increase key
- ⑦ Indicator lamps (MIO, OP1, OP2, AL1, AL2, AU1 and AU2 indicate the I/O actions of the corresponding modules. MAN and PRG are non-applicable for AI-508.)

Basal display status : When power on, the upper display window of the instrument shows the process value (PV), and the lower window shows the setpoint (SV). This status is called basal display status. When the input signal is out of the measurable range (for example, the thermocouple or RTD circuit is break, or input specification sets wrong), the upper display window will alternately display "orAL" and the high limit or the low limit of PV, and the instrument will automatically stop output.



OPERATION DESCRIPTION

- **Set Value Setting:**
In basal display status, if the parameter lock "Loc" is not locked, we can set setpoint (SV) by pressing \leftarrow , \downarrow or \uparrow . Press \downarrow to decrease the value, \uparrow to increase the value, and \leftarrow to move to the digit expected to modify. Keep pressing \downarrow or \uparrow , the speed of decreasing or increasing value gets quick. The maximum value of the setpoint is HIAL.
- **Parameter Setting:**
In basal display status, press ⌚ and hold for about 2 seconds can access Field Parameter Table. Pressing ⌚ can go to the next parameter; pressing \leftarrow , \downarrow or \uparrow can modify a parameter. Press and hold \leftarrow can return to the preceding parameter. Press \leftarrow (don't release) and then press ⌚ simultaneously can escape from the parameter table. The instrument will escape automatically from the parameter table if no key is pressed within 30 seconds.
- **Artificial Intelligence control and auto tuning**
When AI control method is chosen, the control parameters can be obtained by running auto-tuning. At the first time of running auto-tuning, in basic display status, press \leftarrow for 2 seconds, "At" will flash at lower display window and the instrument executes on-off control. After 2 to 3 cycles of on-off action, the instrument will obtain the values of PID control parameters. If you want to escape from auto tuning status, press and hold \leftarrow for about 2 seconds until the "At" flash stops. Depending on the system, the time of auto tuning can be several seconds to several hours. After the auto tuning finishes, the parameter "Ctrl" automatically changes to 3 (the initial value is 1) or 4, and then user can't active auto tuning by pressing \leftarrow for 2 seconds. If user wants to run auto tuning again, setting parameter "Ctrl" to 2. (Refer to the description of parameter "Ctrl")


Note 1: Before running auto-tuning, setpoint should be set to an often-used value or middle value first, and then start auto-tuning. Generally, for application with 10%-80% time on heating can obtain satisfying auto-tuning result. If the heating time is not in the range of 10-80%, the auto-tuning will run for a long time and can't get satisfying result. Raising setpoint for heating time less than 10% or decreasing setpoint for heating time greater than 80% can improve auto-tuning result.

Note 2: Don't modify the setpoint during auto-tuning, it may affect the precision of auto-tuning.

Note 3: The temperature fluctuation caused by interference may affect the precision of auto-tuning. Besides checking the wiring, increasing the parameter dL (input digital filter) can also reduce interference.

Note 4: After auto-tuning finishes, generally, the temperature rising has a little over-shoot (about 1~3℃). It is because considering that the transducer is often placed between the heater and heated material and a little over-shoot can shorten the time of rising to setpoint and save power. (Generally, the temperature of transducer can be raised quick than that of heated material)

PARAMETER AND SETTING

● Press  and hold for 2 seconds to access


Code	Name	Description	Setting Range	Default																								
HIAL	High limit alarm	Alarm on when PV (Process Value) >HIAL; alarm off when PV<HIAL-dF Setting HIAL=9999 can disable high limit alarm.	-1999~ 3000℃	9999																								
HdAL	Deviation high alarm	Alarm on when PV-SV>HdAL; alarm off when PV-SV<HdAL-dF Setting dHAL=999.9℃ can disable deviation high alarm.	0~ 999.9℃	999.9																								
dF	deadband (hysteresis)	Avoid frequent on-off action of alarm or control output because of the fluctuation of PV. “dF” can affect alarm. For example, provided upper alarm parameter "HIAL" is set as 800℃, dF parameter is set as 2.0℃, then: <ul style="list-style-type: none">Instrument is in normal status at the beginning, when process value is greater than 800℃ (HIAL), the upper limit alarm can be triggered.Instrument is in upper alarm status at the beginning, when process value is less than 798℃ (HIAL-dF), then alarm can be cleared. “dF” also affect on-off control and auto tuning. For example, provided SV is set as 700℃, dF parameter is set to 2℃, control is reverse action (heat control). <ul style="list-style-type: none">when process value is greater than 700℃ (SV), then output turn off.when process value is less than 698℃ (SV-dF), output trun on again to start heating. As for on-off control, the larger for dF parameter value, the longer for Ctl (output period time) and the worse for control accuracy. Conversely, the smaller for dF parameter, the shorter for Ctl, then the better for control accuracy, but error action will occur easily due to input fluctuation and shorten the service life of mechanical contactors of relay.	0~ 200.0℃	20℃																								
Ctrl	Control mode	0: on-off control. For situation not requiring high precision; 1: AI control. Allowed to active auto-tuning in basic display status. 2: Run auto-tuning. After auto-tuning finishes, it will automatically change to 3 or 4. 3: AI control, this configuration is automatically set after auto tuning is done. At this setting, starting auto tuning at basic display status is inhibited and it can prevent from running auto tuning by mistake. 4: Similar to Ctrl=3, but parameter P is defined as 10 times as its original value, suitable for rapidly changed temperature (changes by more than 100℃/second). Auto tuning function can automatically set Ctrl to 3 or 4.	0~4	1																								
Ctl	Control period	Small value can improve control accuracy. For SSR or TRIAC output, generally 0.5 to 3 seconds. For Relay output, generally 15 to 45 seconds, because small value will cause the frequent On-Off of mechanical switch and shorten its service life. Ctl is recommended to be 1/4 – 1/10 of derivative time. (It should be integer times of 0.5 second.)	0.5~ 125 seconds	2 seconds																								
Sn	Input specification	<table><tr><td>Sn</td><td>Input spec.</td><td>Sn</td><td>Input spec.</td></tr><tr><td>0</td><td>K</td><td>1</td><td>S</td></tr><tr><td>2</td><td>R</td><td>3</td><td>Spare</td></tr><tr><td>4</td><td>E</td><td>5</td><td>J</td></tr><tr><td>6</td><td>Spare</td><td>7</td><td>N</td></tr><tr><td>8-20</td><td>Spare</td><td>21</td><td>Pt100</td></tr></table>	Sn	Input spec.	Sn	Input spec.	0	K	1	S	2	R	3	Spare	4	E	5	J	6	Spare	7	N	8-20	Spare	21	Pt100	0~21	0
Sn	Input spec.	Sn	Input spec.																									
0	K	1	S																									
2	R	3	Spare																									
4	E	5	J																									
6	Spare	7	N																									
8-20	Spare	21	Pt100																									
Sc	Input offset	Sc is used to shift input to compensate the error caused by transducer, input signal, or auto cold junction compensation of thermocouple. PV_after_compensation=PV_before_compensation + Scb	-199.9~ +400.0 ℃	0.0																								
dL	input digital filter	If measurement input fluctuates due to noise, then digital filter can be used to smooth the input. Parameter “dL” may be configured in the range of 0 to 20, among which, 0 means no filter, 1 means intermediate-value filter, and 2—20 means that intermediate-value filter and second order integral filter working simultaneously. The bigger dL, the more stable the measurement input but the slower the response. Generally if great interference exist, then you can increase parameter “dL” gradually to make momentary fluctuation of measurement input less than 2 to 5 values. If the instrument is being tested at laboratory, then parameter “dL” should be set to 0 or 1 to improve response speed.	0~40	1																								
Loc	Parameter Lock	Loc=0, allowed to modify parameters and setpoint. Loc=1, allowed to view parameters, but can't modify them. And allowed to set setpoint. Loc=2, allowed to view parameters, but not allowed to modify parameters or setpoint (except parameter Loc itself).	0~9999	111																								

Note : Generally, before shipping, the HIAL (high limit alarm) is set to output to AL1, and dHAL (deviation high limit alarm) outputs to AU1 (ALP=5533). However, D2 dimension instruments only have AUX slot. If single relay output module L1 or L2 is installed, HIAL and dHAL should be set to share an alarm output AU1 (ALP=5555); if dual relay output module L5 is installed, HIAL and dHAL should separately output to AU2 and AU1 (ALP=5566). Expert users can change alarm outputs, add low limit or deviation low limit alarm, or hide some parameters by setting internal ALP and EP parameters. You can call your provider for details.


INSTRUMENT INSTALLATION AND WIRING

Wiring graph for instruments with dimension A, E or F


Note: The compensation wires for different kinds of thermocouple are different, and should be directly connect to the terminals. Connecting the common wire between the compensation wire and the terminals will cause measurement error.




Wiring graph for D dimension (72mmX72mm) instruments



Wiring graph for D2 dimension (48X48mm) instruments



Wiring Graph for Thyristor Trigger Output



Note: It is recommended to use the power module, which includes 2 pair of SCRs and diodes. Compared to TRIAC, it is more reliable and consumes less trigger current.